

Sri Vijaya Visakha Milk Producers Company Limited

Akkireddy Palem, BHPV(Post), Vishakhapatnam-530012, Andhra Pradesh

GSTIN 37AAJCS7398P1Z4

TIN/CST NO: 37132885261

Phone : 0891-2517230

Excise Regn.No: AAJCS7398PEM001

Fax : 0891-2757708

CIN : U15209AP2006PTC048708

Email : purchase@visakhadairy.com

Vendor Code: V02116

PURCHASE ENQUIRY

ENQUIRY No : PEQE5-25-226 Date : 04-11-2025

Indent No: S1-25-84 Indent Dt: 27-10-2025

Unit-Name RANGAMPETA

Department : E & M

Center-Name: E&M - Boiler

Bid Date 10-11-2025

Ph.

Email ID :

Please supply the Equipment/Material Specified below. All terms and conditions incorporated here in on the reverse / attached to this purpose order including general terms and conditions of purchase.

ITEM ANNEXURE ATTACHED

Delivery Schedule : immediately

Place of Delivery : Visakha dairy -Visakhap

Special Note : Please quote your offer as per enquiry specifi

Narration :

Delivery Terms :

Note : Please send your offer with GST Details

For VISAKHA DAIRY

Sr. Manager (Purchase)

Sri Vijaya Visakha Milk Producers Company Limited

Akkireddy Palem, BHPV(Post), Vishakhapatnam-530012, Andhra Pradesh

GSTIN 37AAJCS7398P1Z4

TIN/CST NO: 37132885261

Phone : 0891-2517230

Excise Regn.No: AAJCS7398PEM001

Fax : 0891-2757708

CIN : U15209AP2006PTC048708

Email : purchase@visakhadairy.com

PURCHASE ENQUIRY

P. No : PEQE5-25-226

Date : 04-11-2025

ITEM ANNEXURE FOR ABOVE PURCHASE Enquiry

S.No	Code	Description	Toler%	Units	Indent No:	Quantity
1	MSIMSP899	Digital Water Level Recorder (Piezometer) With Telemetry Remarks:REQUIRED PIEZOMETER WITH INSTALLATION Specifications:Copper Cable length:100 meters (Approx.) Make: Engineering and Electronics Model No: EE-GWLR-01 Data Storage Capacity: 5 years Logging Interval: per minute Also fulfill all technical specifications of CGWA standards Communication: Inbuilt Dual SIM based 4G/2G communication Parameters to monitor: The meter should transmit the fluctuations in ground water levels and quality and securely maintained in cloud Data Management Software: The data shall be acquired in a secured cloud and shall provide a web dashboard and mobile application to view and download data of the DWLR.		Nos	S1-25-84 27-10-2025	1.00
2	MSIMSP891	Tamper Proof Digital Flow Meter With Telemetry,Size: 50 Mm Remarks:REQUIRED FLOWMETER WITH INSTALLATION Specifications:Metering Technology: Ultrasonic, Max. Flow Rate: 25m3/hr.ISO 4064: 2014 Standard, Shall have IP68 ingress protection and fulfill CGWA Standards Communication: Inbuilt Dual SIM based 4G/2G communication, Power Source: The meter and telemetry shall be operated using non-rechargeable batteries with minimum 3 years battery life with minimum 3 transmission per day		Nos	S1-25-84 27-10-2025	1.00

Sr. Manager (Purchase)